الآيـــة

بِسُمِ ٱللَّهِ ٱلرَّحْمَنِ ٱلرَّحِيمِ

اللَّهِ ٱلرَّحْمَنِ ٱلرَّحِيمِ

اللَّهِ وَرَبِّكَ ٱلْكَارَمُ اللَّهِ ٱلَّذِي خَلَقَ اللَّهِ خَلَقَ ٱلْإِنسَيْنَ مِنْ عَلَقٍ اللَّهِ اللَّهِ اللَّهِ عَلَمَ اللَّهُ عَلَى اللَّهُ عَلَمُ اللَّهُ عَلَمَ اللَّهُ عَلَمُ اللَّهُ عَلَى اللَّهُ عَلَى اللَّهُ عَلَى اللَّهُ عَلَى اللَّهُ عَلَمُ اللَّهُ عَلَى اللَّهُ عَلَمُ اللَّهُ عَلَى اللَّهُ عَلَمُ اللَّهُ عَلَى اللْعَلَمُ عَلَى اللَّهُ عَلَى الللَّهُ عَلَى الْعَلَمُ عَلَى اللَّهُ عَلَى اللَّهُ عَلَى الْعَلَمُ عَلَى اللَّهُ عَلَى اللْعَلَمُ عَلَى الْعَلَمُ عَلَى الْعَلَمُ عَلَى اللْعَلَمُ عَلَى الْعَلَمُ عَلَى اللْعَلَمُ عَلَى ا

لَـمْ يَعْلَمُ ۞

صدق الله العظيم سورة العلق

Dedication

This thesis is dedicated to:

Soul my father

My mother

My brothers

My sisters

My family

My teachers

And to my friends

Acknowledgment

My faithful thanks and praise to Allah for providing me with health and strength to conduct this study. I would like to express my deep gratitude with special respect to my Supervisor *Dr. Hussein Ahmed Hassan* who gave me much of his time for suggestion and careful supervision during this period, to product this study. I wish to extend my thanks to *Private Medical Institution*. Lastly, Iam also so grateful for *Mr. OmerMohammed ElhadiBabiker* and to every person helped me in gathering information and guiding me in making this study.

Abstract

The aim of this study is to evaluate the of beam alignment of x-ray tube accuracy in the Khartoum hospitals, this study include 12 x-ray machines in 12 hospitals.the accuracy of radiation was tested and the most important results obtained that the alignment is 91.7% & misalignment is just 8.3% in along cassette, and 75% in across cassette 25% misalignment . in the central ray the alignment reached 83.3 % & misalignment 16.7%.

This study confirms that x-ray machine in the Khartoum state hospitals operate at a good level.

الملخص

الهدف من هذه الدراسة تقييم دقة تطابق شعاع انبوب الاشعة السينية في مستشفيات و لايةالخرطوم شملت هذه الدراسة 12 جهاز اشعة سينية في 12 مستشفى. تم اختبار دقة تطابق الشعاع.

وقد وجد ان نسبة التطابق وصلت الى 91.7 بالمائة ونسبة عدم التطابق 8.3 بالمائة على طول الفيلم, وعلى عرض الفيلم قد بلغت نسبة التطابق 15 بالمائة, اما في وسط الفيلم بلغت نسبة التطابق 83.5 بالمائة وعدم التطابق 16.7 بالمائة.

تؤكد هذه الدراسة ان اجهزة الاشعة السينية في مستشفيات ولاية الخرطوم تعمل بمستوى جيد.

LIST OF CONTENTS

NoItems	Page No	
الآية	I	
Dedication	II	
Acknowledgment	III	
Abstract (English)	IV	
Abstract (Arabic)	V	
Contents	VI	
Chapter One		
1.1 introduction	1	
1.2 problem of the study	2	
1.3 objective of the study	2	
1.3.1 general objective	2	
1.3.2 specific objective	2	
1.4 lay out	3	
Chapter Two		
2.1 Background and literature Review	4	
2.2 X-ray Tubes	4	
2.2.1 Cathode	6	
2.2.2 Anode	7	
2.3 Production of x-rays	8	
2.3.1 Bremsstrahlung Spectrum	8	
2.3.2 Characteristic x-ray Spectrum	12	
2.4. Interactions of X-ray with matter	14	
2.4.1 Rayleigh Scattering	14	
2.4.2 photoelectric effect	15	
2.4.3 compton scattering	17	
2.4.4 pair production	20	
2.5. The Quality Control Program	20	
2.5.1. Equipment Selection	20	
2.5.2. Acceptance Test	21	
2.6. Quality Control	21	
2.6.1. Documentation	22	
2.6.2. Staffing Considerations	22	
2.6.3. QC INSTRUMENTATION	22	

2.7. Light field/X-ray field alignment (congruence)	23			
2.7.1. Method of test	23			
2.7.1.1 Light field /beam alignment test	23			
2.7.1.2 Equipment	24			
2.7.1.3 Procedure	24s			
2.7.2 Test Procedure - Routine Test 25				
2.7.3. Alternate Method	26			
2.8. Previous studies	28			
Chapter Three				
3. Materials and Methods	32			
3.1 Materials	32			
3.1.1 Study sample	32			
3.1.2 Equipment	32			
3.1.3 Procedure	34			
3.2. Acceptance Parameters	34			
3.3 Record Keeping	34			
3.2.2 Analysis	34			
Chapter Four				
4. Results	37			
Chapter Five				
5.1 Discussion				
5.2 Conclusion 47				
5.3 Recommendation 48				
5.4 References 4				

List of Tables

Table	Item	Page number
4.1	distribution of machines according to	37
	manufactures	
4.2	total misalignments measurements	38
4.3	Total measurements of alignment of X-ray field	40
	to cassette holder	
4.4	show beam alignments accuracy in total along	41
	cassette	
4.5	show beam alignments accuracy in total across	42
	cassette	
4.6	show accuracy of alignment of X-ray field to	43
	cassette holder	

List of Figures

Figures	Item	Page
		number
2.1	Minimum requirements for x-ray production include a	9
	source and target of electrons, an evacuated envelope,	
	and connection of the electrodes to a high-voltage source	
2.2	Bremsstrahlung radiation arises from energetic electron	10
	interactions with an atomic nucleus of the target material	
2.3	Generation of a characteristic x-ray	13



2.4	Generation of a characteristic x-ray	13
2.5	Rayleigh scattering	15
2.6	Compton scattering	16
2.7	show photoelectric absorption	18
2.8	Beam alignment test tool	24
2.9	Template image	25
2.10	showstheArrangementofexperimental set up for exposure	26
2.11	shows the arrangement of experimental set for exposure	27
3.1	Diagram of the layout for the nine penny test	33
3.2	Image of the nine penny test	33
3.3	Determining the Total Misalignment of the Light Field	35
	and the X-ray Field	
3.4	Determining Alignment of X-ray Field to Cassette	36
	Holder	
4.1	distribution of machines according to manufactures	38
4.2	Total misalignments measurements for along cassette &	39
	across cassette	
4.3	Total measurements of alignment of X-ray field to	40
	cassette holder	
4.4	show beam alignments accuracy in total along cassette	41
4.5	show beam alignments accuracy in total across cassette	42
4.6	show accuracy of alignment of X-ray field to cassette	43
	holder	

List of abbreviation

QA: Quality Assurance

QC: Quality Control

SID: Source to image distance

LBDs: light beam diaphragms

FFD: focal film distance

AAPM: American Association of physicists in Medicine

AL1 + **AL2**: Total along cassette misalignment in cm

AC1 + AC2: Total across cassette misalignment in cm

CFR: Code of Federal Regulation

mAs: milliAmbere second

KVp: Kilovolt peak

BLD: Beam lighting device

CR: Computed Radiographer

NCRP: Nuclear commission radiation protection